Cocaine and Metabolites (TMS Derivatives) on Rxi®-5Sil MS (100 ng/mL)



GC_CF1149

Peaks	m/z 1	m/z 2	m/z 3
1. Ecgonine methyl ester	82	96	256
2. Ecgonine	82	96	356
3. Cocaine	82	182	303
4. Cocaethylene	82	196	317
5. Benzoylecgonine	82	96	346
6. Cannabinol*	409	--	--

Column	Rxi®-5Sil MS, 30 m, 0.25 mm ID, 0.25 µm (cat.# 13623)
Sample	
Diluent:	Butyl chloride
Conc.:	100 ng/mL
Injection	
Inj. Vol.:	1 µL splitless (hold 1 min.)
Liner:	Single Gooseneck w/Wool (cat.# 22286-200.1)
Inj. Temp.:	250 °C
Purge Flow:	20 mL/min.
Oven	
Oven Temp:	100 °C to 200 °C at 30 °C/min. to 300 °C at 15 °C/min.
Carrier Gas	He, constant linear velocity
Linear Velocity:	40 cm/sec., 12.5 psi, 86.2kPa @ 100 °C
Detector	MS
Mode:	SIM
Transfer Line Temp.:	310 °C
Source Temp.:	250 °C
Solvent Delay Time:	4 min.
Tune Type:	PFTBA
Ionization Mode:	EI
Instrument	Shimadzu 2010 GC & QP2010+ MS
Notes	Samples were prepared as follows: Standards brought to dryness under nitrogen, then 50 µL BSTFA + 1%TMCS (cat.# 35606) added. 50 µL pyridine was then added, and samples were incubated at 70°C for 30 min. After incubation, samples were diluted with butyl chloride. * Used as derivitazation check

Restek Corporation 110 Benner Circle Bellefonte, PA 16823
814-353-1300 • 800-356-1688 • Fax: 814-353-1309 • www.restek.com